

Telephone: +44 (0)20 8805 3424
Fax No: +44 (0)20 8805 3054

Leaside Industrial Estate
Stockingswater Lane
Enfield, Middx. England
EN3 7PH

DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;
IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993
and thus meet **THE LOW VOLTAGE DIRECTIVE (LVD)**
73/23/EEC and amending directive 93/68/EEC

PART No:

129303/OS

PRODUCT TYPE / DESCRIPTION:

OEM Fixed Lead Prod to Shutter Plug Test Lead Set

IEC 1010 RATING:

1000V CAT III (3)

MAXIMUM RATED CURRENT:

20A

POLLUTION DEGREE:

2

INCLUDES THE FOLLOWING ASSEMBLIES / COMPONENTS SUBECT TO INDIVIDUAL TEST.

DESCRIPTION:

OEM Fixed Lead Prod

Shutter Plug

1mm sq D/S Cable


PART No:

125074

125106

119054

Approved Signatories: J.DEVEREUX () A.WILLIAMS (✓) V.NEGLIA ()
Signed on behalf of **SILVERTRONIC Ltd.**





DATE: 25th February 2004



All Testing carried out by Third Party Testhouse
CRANAGE EMC TESTING Ltd.
(LVD TESTING DEPARTMENT)

| | | | | | | | | | | | | |
|---|----------|-----------|---|-----------|----------|-------------|--|--|--|--|--|-------------|
| SILVERTRONIC LTD. | Date | Drn.By | Title OEM FLP TO OS PLUG LEAD SET | ISSUE No. | 1 | | | | | | | Part No. |
| | 04/03/04 | Vince | | | 04/03/04 | PROD. ISSUE | | | | | | 129303/OS |
| This drawing is not to be disclosed to a third party without the authority from the company.© | | Scale | IEC 1010 COMPLIANT RATING: 1000 VOLTS CAT III 20 AMPS POL DEG. 2 | | | | | | | | | PROJECT No. |
| | | FULL SIZE | | | | | | | | | | |



| | | | | |
|---|-------------------------------|-------------|------------------------------------|--|
| ASSEMBLY JIG NUMBERS 125074= J164 or 167 + 166 | PACKAGING POLY BAG: 111500 | NO PRINTING | TOLERANCES CABLE LENGTHS +/- 5% | DO NOT SCALE. ALL DIMENSIONS IN mm. |
|---|-------------------------------|-------------|------------------------------------|--|