



Description

The Vermason Slim Earth Bonding Point Bar (or Common Ground Point) has been designed to be installed underneath bench tops where it will be accessible to all connecting ground cords and where it is unlikely to be knocked and damaged or hinder the operator.

With a width of only 15mm it is especially suitable for benches with narrow edges. The product is supplied with reverse-mounted screws held in position by self-retaining washers, thereby making the installation process simple. The distance between screw centres is 89mm. The design of the ABS plastic moulding provides strength to the assembly and the rounded corners ensure comfort during use for the operator.

Connect earthing cord (3m long) to a earth protective ground. A spare ring terminal is supplied if the cord needs to be shortened.

Resistors are built-in to each connector. The normal earth contact is linked via individual 1 megohms resistor rated at 1/2W. In the event of a workstation element becoming live the resistor will limit the fault current to less than 0.3mA at 240V. The resistor is capable of withstanding 500VDC for 6 minutes.

Per EN 61340-5-2 paragraph 5.1.1 "A nominal 1 megohms resistor is commonly used in wrist straps and to ground work surfaces. In the event of an operator touching an energized conductor, for the normal mains electricity supplies this resistor will limit the current flowing through the person to less than 0.5mA. Current limiting or cut-off devices should also be considered and in some areas may be demanded by legislation."

The operator connects the wriststrap to the male snap stud via a coiled cord. A second stud can be used to connect a working surface. The third connection is available for guest use or to ground another item such as a trolley. Periodic verification of the ground path should be part of the ESD control program.

Item #	Description
VER-26989	EBP, SLIM BAR, 3M CORD, 3x10MM STUDS

"The protection of ESDS [ESD sensitive item] is accomplished by providing a ground path to bring ESD protective materials and personnel to the same electrical potential. All conductor and dissipative items in the environment, including personnel, shall be bonded or electrically connected to a known ground or common connection point [Earth Bonding Point]." [CLC/TR 61340-5-2 User guide clause 4.4.1 Grounding/bonding systems Introduction]

Specifications and procedures subject to change without notice.



Made in the United Kingdom

Vermason

Slim EBP Bar

VERMASON
UNIT C, 4TH DIMENSION, FOURTH AVENUE, LETCHWORTH,
HERTS, SG6 2TD UK
PHONE: +44 (0) 1462-672005
E-MAIL: Service@Vermason.co.uk, INTERNET: Vermason.co.uk

Drawing Number
VER-26989

DATE:
September
2015