Black: 76-1672 Red: 76-1674

most famous

manufacturers.

DATA SHEET (page 1 of 2).

Designation: 4 mm Banana (female) Jack (socket) w/ Solder Wire Attachment and M6 Threaded Stud and Hex Nuts. Quick Radial Wire Attachment (2 mm diameter Hole).

Blue: 76-1476 Applications: repairing or making of panels or boxes providing quick wire attachments How to implement: and 4 mm banana connections for power supplies, measurements, controls, tests, ... Step 1 of 5. I gather an open-end spanner SW8 mm, an insulating panel with the specifications below, and a tool to drill the insulating panel as below. Thanks to the nuts, the Insulating panel: socket can be removed 17 mm when the button from the panel to be ← Maxi. thickness: 11 mm is fully unscrewed 18 replaced or re-used. (to comply with a 1 mm-thick lug). Pitch circle diameter to drill the insulating panel : $\emptyset 6.0 (+0.1/-0)$ mm IP2X touchproof I drill the insulating panel as above with the tool. protection when the button is Step 3 of 5. If the nuts and the star washer are mounted on the socket then I remove screwed fully. them. I tighten the button. I push the socket into the hole of the insulating Ø3.5 Ø2 Ø3.5 panel as shown below. Ø2 mm hole. Ø7.5 $M6 \times 0.75$ Ø10.8 (thickness 2 mm.) Insulating panel. Star washer. Nuts. 14 mm when the button The button can is fully screwed. not be removed How to use the wire attachment: because it is Front face of the panel. Rear face of the panel. designed to be I gather a stranded or solid wire unloosable. Step 4 of 5. I put the star washer on the rear side of the socket. Then I put one of the two with the specifications below nuts on the rear side of the socket too. I hold the button, I hold the nut with and a tool to strip the wire. the spanner SW8, then I screw and tighten it (2.3 N.m maxi. torque). Now the socket is attached to the insulating panel as shown below. I strip the end of the wire on 10 mm at least. Ring lug (not provided) I unscrew the button. I insert the wire into the radial Ø2 mm hole. I screw and tighten the button. The 4 mm banana The screwing The terminal complies with usual female connection button 4 mm² - 6 mm² ring lug tightened Step 5 of 5. To connect the socket I can solder (150 Watt maxi. iron solder with lead-tin 10 mm mini... complies with the nontightens the between the two nuts. It offers or lead-free tin) a wire on its terminal so I screw and tighten the other nut shrouded 4 mm banana wire inserted solder wire attachment too with first (2.3 N.m maxi. torque) (it is a lock nut). Or I can attach a wire plugs of the worldwide into the radial lead-tin or lead-free tin and terminated by a usual ring lug (4 mm² - 6 mm² ring lug) so I put the ring

1.50 mm² maxi.

(approx. AWG16).

150 W maximum soldering iron.

Ø2 mm hole.

torque). Then the socket is ready to use.

lug (picture above) then I screw and tighten the other nut (2.3 N.m maxi.



DATA SHEET (page 2 of 2).

ACCESSIBLE. Able to be touched with a standard test finger or test pin.

GLOSSARY:

BASIC INSULATION. Insulation of HAZARDOUS LIVE parts which provides basic protection.

CAT II. Measurement or overvoltage category II. For measurement performed on / equipment connected to the building wiring.

CAT III. Measurement or overvoltage category III. For measurement performed on / equipment connected to part of a building wiring installation.

CAT IV. Measurement or overvoltage category IV. For measurement performed on / equipment connected to the origin of the electrical supply to a building.

CLEARANCE. Shortest distance in air between two conductive parts.

CREEPAGE DISTANCE. Shortest distance along the surface of a solid insulating material between two conductive parts.

CTI. Comparative Tracking Index of the insulating material in accordance

DOUBLE INSULATION. Insulation comprising both BASIC INSULATION and SUPPLEMENTARY INSULATION.

that occasionally a temporary conductivity caused by condensation is expected.

REINFORCED INSULATION. Insulation which provides protection against electric shock not less than that provided by DOUBLE INSULATION.

"RoHS". European Directive 2011/65/EU on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

SOLID INSULATION. Insulating materials.

SUPPLEMENTARY INSULATION. Independent insulation applied in addition to BASIC INSULATION in order to provide protection against electric shock in the event of a failure of BASIC INSULATION.

TRANSIENT OVERVOLTAGE. Short duration overvoltage of a few milliseconds or less, oscillatory or non-oscillatory, usually highly damped.

WORKING VOLTAGE. Highest r.m.s. value of the a.c. or d.c. voltage across any particular insulation which can occur when the equipment is

Black : 76-1672 Red: 76-1674 Blue: 76-1476

Designation: 4 mm Banana (female) Jack (socket) w/ Solder Wire Attachment and M6 Threaded Stud and Hex Nuts. Quick Radial Wire Attachment (2 mm diameter Hole).



Electrical safety 33 V AC / 70 V DC	Very low voltages only: 33 V AC / 70 V DC,	EN / IEC 60529. European / international standard regarding the degrees of protection provided by enclosures.
	36 A (at +40 °C).	EN / IEC 61010-1. European / international standard regarding the safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements.
Operating temperature range Conformity	 -20 °C mini., +80 °C maxi. (please see above too). European Directive "RoHS" 2011/65/EU. European REACH regulation n°1907 / 2006. 	EN / IEC 61010-031. European / international standard regarding the safety requirements for electrical equipment for measurement, control and laboratory use – Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test.
Environment	 "RoHS" compliant, Pb ≤ 4 % in conductor, Pb ≤ 0.1 % in insulator, Hg ≤ 0.1 %, Cr VI ≤ 0.1 %, Cd ≤ 0.01 %, PBB ≤ 0.1 %, and PBDE ≤ 0.1 %. REACH compliant, no substances from the candidate list of SVHC for authorisation at mass concentrations greater than 0.1 % 	"LVD". European Directive 2014/35/EU on the harmonization of the laws of Member States relating to electrical equipment designed for use within certain voltage limits. (Usually called the Low Voltage Directive.) MAINS. Low-voltage electricity supply system to which the equipment concerned is designed to be connected for the purpose of powering the equipment.
Materials	Conductors : nickel-coated (or gold-coated) brass and steel. Insulators : please contact us.	MAINS CIRCUIT. Circuit which is intended to be directly connected to the MAINS for the purpose of powering the equipment.
Colors	Black Red Yellow Green Blue White	OVERVOLTAGE CATEGORY. Numeral defining a TRANSIENT OVERVOLTAGE condition.
Weight	0.007 kg.	POLLUTION. Addition of foreign matter, solid, liquid or gaseous (ionized gases), that may produce a reduction of dielectric strength or surface
Origin	Designed and manufactured in France.	resistivity. POLLUTION DEGREE. Numeral indicating the level of POLLUTION that may be present in the environment.
Reliability benchmark	Year of 1st placing on the market 1980.	
Packaging	One piece per bag (in one bag : 1 socket + 1 starwasher + 2 nuts).	POLLUTION DEGREE 1. No POLLUTION or only dry, non-conductive POLLUTION occurs, which has no influence. POLLUTION DEGREE 2. Only non-conductive POLLUTION occurs except that consciously a transport mendantistic according to the conductive pollution.