

3421594

ALL RIGHTS RESERVED. NO PORTION OF THIS PUBLICATION, WHETHER IN WHOLE OR IN PART CAN BE REPRODUCED WITHOUT THE EXPRESS WRITTEN CONSENT OF SPC TECHNOLOGY.

REVISIONS			DOC. NO. SPC-F004 * Effective: 12/21/98 * DCP No: 690					
DCP #	REV	DESCRIPTION	DRAWN	DATE	CHECKD	DATE	APPRVD	DATE
		NOT Released	JWM	10/9/98				



General purpose replacement probe with 3-position switchable attenuation, x1, x10 and Ref. 100MHz @ -3dB x 10 position, 10MΩ impedance (1MΩ scope input), 18pF capacitance, <3.5ns risetime. 6MHz @ -3dB x 1 position, 1MΩ impedance (scope input), 90pF capacitance, 58ns risetime. Ref position disconnects the probe tip and grounds the scope input. 1.2m cable length with low-noise compensation, 10-35pF, in BNC box.

SPC-F004.DWG

**DISCLAIMER:**  
ALL STATEMENTS AND TECHNICAL INFORMATION CONTAINED HEREIN ARE BASED UPON INFORMATION AND/OR TESTS WE BELIEVE TO BE ACCURATE AND RELIABLE. SINCE CONDITIONS OF USE ARE BEYOND OUR CONTROL, THE USER SHALL DETERMINE THE SUITABILITY OF THE PRODUCT FOR THE INTENDED USE AND ASSUME ALL RISK AND LIABILITY WHATSOEVER IN CONNECTION THEREWITH.



UNLESS OTHERWISE SPECIFIED, DIMENSIONS ARE FOR REFERENCE PURPOSES ONLY.	DRAWN BY:	DATE:	DRAWING TITLE:			
	Jeff McVicker	10/9/98	100 MHz Oscilloscope Probe			
	CHECKED BY:	DATE:	SIZE	DWG. NO.	ELECTRONIC FILE	REV
			A	76-102	84N2144.dwg	
	APPROVED BY:	DATE:	SCALE: NTS	U.O.M.: INCHES [mm]	SHEET: 1 OF 1	