



# DECLARATION OF CONFORMITY

Apparatus Identification

**Fluke Model 179, 177, 175 TRMS Multimeter**

Apparatus Classification

**Test and Measurement Equipment**

Statement of Conformity

Based on sample product test results using appropriate standards, and in accordance with the following EC Directives, the Fluke Corporation hereby declares the **179, 177, & 175 TRMS Multimeter** to be in conformity with:

**EC Directive 89/336/EEC, Electromagnetic Compatibility, and  
EC Directive 73/23/EEC, Low Voltage (Amended 93/68/EEC).**

Sample Product Testing for EMC

**Tested By** Northwest EMC (with input from Fluke Corporation)  
22975 NW Evergreen Pkwy Ste. 400  
Hillsboro, OR 97124

**Standards Used** **EN61326-1**, (1998) Electrical equipment for measurement, control and laboratory use.

**Report ID** FLUK0019, issued 07-September-2000,  
for **Fluke Models 179, 177, 175 TRMS Multimeter.**

Sample Product Testing for Safety (applicable only to Ext Power Supply accessory)

**Tested By** **Fluke Corporation**  
P.O. Box 9090  
Everett, WA 98206-9090

**Standards Used** **EN 61010-1:1993+Amendment 1:1992, Amendment 2:1995 + EN 61010-1 2<sup>nd</sup> Edition**

**Report ID** Product Safety Conformance Verification Report for the **Fluke 179, 177, 175 TRMS Multimeter** issued 07-November-2000.

Manufacturer

**FLUKE Corporation**  
6920 Seaway Boulevard  
PO Box 9090  
Everett, WA 98206-9090  
United States of America

EU Representative

**FLUKE Europe B.V.**  
Science Park Eindhoven 510  
5692 EC SON Eindhoven  
The Netherlands

  
James Cavoretto, Vice President  
Fluke Industrial Group

  
Jan Roetman, General Manager  
Fluke Europe B.V.