ROHDE&SCHWARZ





This is to certify that:

Equipment Type Stock No. Designation

FPC1000 1328.6660n02 SPECTRUM ANALYZER FPC1500 1328.6660n03 SPECTRUM ANALYZER

The object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

Electrical Equipment (Safety) Regulations 2016 (S.I. 2016/1101)

Harmonized standards applied: EN 61010-1:2010/A1:2019

Electromagnetic Compatibility Regulations 2016 (S.I. 2016/1091)

Harmonized standards applied: EN 61326-1:2013

EN 55011:2016 + A1:2017 + A11:2020

Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (S.I. 2012/3032)

Harmonized standards applied: EN IEC 63000:2018

The immunity test requirements for equipment intended to be used in industrial electromagnetic environment (acc. to EN 61326-1, Table 2) have been used as a basis.

The limits of radio interference for Group I, Class A equipment (acc. to EN 55011) have been used as a basis.

ROHDE & SCHWARZ GmbH & Co. KG Mühldorfstr. 15, D-81671 Munich

Corporate Quality Management GF-QP / Robert Lambertus EN MH Page 1 / 1