Features

Regulated Converter

- 400/450 Watt convection cooled (115/230VAC)
- 600 Watt forced air or peak power
- 5VSB Output
- Redundant operation; active current sharing
- Remote sensing, CTRL ON/OFF, PMBus™
- IEC60601-1 2x MOPP insulation, BF-ready

Description

RACM600-L/OF Series AC/DC power supply units are designed for operation in natural convection and in systems with certain airflow ventilation to deliver 400 to 600Watt output power. Safety approvals to Medical IEC 60601-1-2 and to IT and industrial IEC 62368 standards and operation with worldwide input voltage conditions from 80 to 275Vac in altitudes up to 5000m make these chassis mount units ideal for global use in medical, industrial or IT related automation processes. For enhanced reliability requirements of applications redundant operation is supported with active current sharing. An additional 5V Standby output powers housekeeping circuitry to control remote on/off and monitoring functions which are available via PMBus™ I²C interface. EN55032 class "B" EMC compliance is achieved without any external components which underlines the versatility of these power supplies.

Selection Guid	e				
Part Number	Input Voltage Range [VAC]	Nom. Output Voltage [VDC]	Max. Output Current [A]	Max. Output Power [W]	Efficiency typ. ⁽¹⁾ [%]
RACM600-12SL/OF	80-275	12	50	600	92
RACM600-24SL/0F	80-275	24	25	600	93
RACM600-48SL/0F	80-275	48	12.5	600	93

Notes:

Note1: Efficiency is tested at 230VAC and full load at +25°C ambient

Model Numbering



Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

BASIC CHARACTERISTICS

Parameter	Condition	Min.	Тур.	Max.
Nominal Input Voltage	50/60Hz	100VAC		240VAC
Operating Range (2, 3)	47-63Hz	80VAC		275VAC
Operating hange ***	DC	120VDC		300VDC
Input Current	80VAC			9A
input Guirent	120VDC			5.7A
Inrush Current	cold start at 25°C			20A
Input Frequency Range	AC Input	47Hz		63Hz
Minimum Load		0%		
Power Factor	EN61000-3-2, Class A compliant		0.9	
Start-up Time	MAIN ON			2.5s
Start-up Tillle	CTRL ON			150ms
Rise Time				150ms
Hold-up Time			20ms	
Periodic and Random Deviation (PARD)	20MHz BW, 10µF Tan. and 1µF MLCC			1%p-p

Notes:

Note2: The products were submitted for safety files at AC and DC-Input operation.

Note3: Refer to "Rating Graphs of continuous Operation"

continued on next page



RACM600-L

600 Watt 7.7" x 4" Open Frame Single Output













IEC/EN62368-1 certified
UL62368-1 certified
CSA/CAN C22.2 No. 62368-1 certified
ANSI/AAMI ES60601-1 certified
CSA/CAN C22.2 No. 60601-1:14 certified
IEC/EN60601-1 certified
EN55032 compliant
EN55024 compliant
EN60601-1-2 compliant
CB Report



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)






ADDITIONAL FEATURES						
Parameter		Conc	lition	Min.	Тур.	Max.
ON/OFF CTRL	CON3		MAIN OUTPUT ON			open
(logic can be switched with PMBus™)	CONS		MAIN OUTPUT OFF	CTRL (pin10) shorted to 5VSI	B_RTN (pin3,4,7)
	on board notil refer to "O	utnut	V _{out} = 12VDC	9.6VDC		14.4VDC
Output Voltage Adjustability (4)	on-board poti, refer to "O Current vs. Output Voli	•	V _{OUT} = 24VDC	19.2VDC		28.8VDC
	Gurrent vs. Gutput von	laye	V _{OUT} = 48VDC	38.4VDC		56VDC
Remote Sense (5)	total voltage drop compens	sation f	or +Sense and -Sense connection			200mV
Power OK LED		LED =	green	turn ON as soon a	s PSU_GOOD Sig	gnal is set to high

Notes:

Note4: By trimming up, decrease output power. By trimming down, do not exceed maximum continuous output current Note5: Do not short or reversely connect +Sense to -Sense, this can cause damage to the supply

Output Current vs. Output Voltage



5VSB OUTPUT (6)					
Parameter	Condition	Min.	Тур.	Max.	
Nominal Output Voltage				5VDC	
Max. Output Current				500mA	
Max. Output Power				2.5W	
Max. Capacitive Load				1000μF	
Over Voltage Protection (OVP)			5.5-6VDC, latch off		
Over Current Protection (OCP)	of rated I _{OUT}		1-1.3A, auto recovery		
Short Circuit Protection (SCP)				auto recovery	
Over Temperature Protection (OTP)				auto recovery	

Notes:

Note6: There is no galvanic isolation between AUX GND and Main Output GND. Regulations for 5VSB Output are stated under "REGULATIONS"



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

REGULATIONS		
Parameter	Condition	Value
Output Accuracy (MAIN and 5VSB output)		±2.25% max.
Line Regulation (MAIN and 5VSB output)	low line to high line, full load	±0.25% typ.
Load Regulation (MAIN and 5VSB output)	0% to 100% load	1.0% typ.
Dynamic Load Regulation	50% step from 5% load (1A/μs), tested with 10μF Tan. and 1μF MLCC	5.0% max.

PROTECTIONS			
Parameter	Тур	De .	Value
Internal Input Fuse	DC input compli	ant, dual-fusing	2x T10A
Short Circuit Protection (SCP)			hiccup, auto recovery
Over Voltage Protection (OVP)	$V_{\text{OUT}} = 1$ $V_{\text{OUT}} = 2$	24VDC	15VDC - 17.5VDC, latch off 30VDC - 35VDC, latch off
Over Voltage Category (OVC)	V _{OUT} = 4	18ADC	58.5VDC - 63VDC, latch off OVCII
Over Current Protection (OCP)	of rate	ed l _{out}	108-140%, auto recovery
Over Temperature Protection (OTP)			auto recovery
Isolation Voltage (safety certified) (7)	I/P to O/P (reinforced) I/P and O/P to Case (basic)	1 minute	4kVAC (2MOPP) 1.5kVAC (1MOPP)
Insulation Grade			reinforced
	1 1 100/40 00/1	Normal condition	150µA max.
	low line 132VAC , 63Hz	Single Fault	250µA max.
Leakage Current Input to Earth GND	1: 1: 004/40 0011	Normal condition	300µA max.
	high line 264VAC, 60Hz	Single Fault	500μA max.
		Normal condition	60μA max.
Lastrana Oromant Ortant to Fauth OND	00.41/4.0	Single Fault (neutral open)	80µA max.
Leakage Current Output to Earth GND	264VAC , 63Hz	Single Fault (ground open)	150µA max.
		AC Back-drive Fault	550μA max.
Class of Equipment			Class I
Medical Device Classification	according to	IEC60601-1	designed to support Type BF applied part
Note	es:		

Note7:	For repeat Hi-Pot testing, reduce the time and/or the test voltage

ENVIRONMENTAL					
Parameter	Condition		Value		
Operating Temperature Range	refer to "Rating Graphs of continuous Operation"	T _{BASE} temperature	-20°C to +70°C		
Operating Altitude (8)	according to 62368-1				
Operating / untade	according to 60601-1		3000m		
Operating Humidity	non-condensing	non-condensing			
Pollution Degree					
Vibration (non-operating)	2.09Gr.m.s., 5Hz to 500Hz, 20 minutes per side (3 planes)		according to IEC 60068-2-6		
Shock (non-operating)	50G, 11ms, 3 shocks for each direction	50G, 11ms, 3 shocks for each direction			
MTBF	according to Telcordia SR-332, Issue 3, 25°C ambient, 9	90% confidence level	500 x 10 ³ hours		
Design Lifetime (capacitor)	nom. Vin, 80% load, 45°C ambient		87.6 x 10 ³ hours		

Notes:

Note8: Recognized by safety agency for safe operation up to 5000m. High altitude operation may impact the performance and lifetime Ambient temperature decreases by 1°C per 305m altitude increase


continued on next page

www.recom-power.com REV.: 1/2022 PA-3



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)



PEAK LOAD CAPABILITY OF 12V & 48V MODELS (not applicable for 24V model)

Guideline for sporadically occurring peak loads:

800 Watt max. @ 40°C ambient with a maximum duty cycle of 0.5% .. for still air convection

4 % ... at 1.5m/s provided system airflow

continued on next page



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

PEAK LOAD CAPABILITY OF 12V & 48V MODELS (not applicable for 24V model)

Calculation Guideline for recurrent dynamic load cycles:

Peak Load calculation for recurrent dynamic loading			alculation for onvection	Peak Load Capability with 1.5m/s airflow
Parameters	Units	RACM600-12SL/0F	RACM600-48SL/0F	RACM600-12SL/0F RACM600-48SL/0F
P _{nom} = rated output power	[W]	≤320	≤430	≤600
P _P = peak output power	[W]	≤800	≤800	≤800
P _r = recovery output power	[W]			
t ₁ = peak time set	[s]	≤5s	≤5s	≤10s
t ₂ = recovery time	[s]	min. 10 x t ₁	min. 8 x t ₁	min. 3 x t ₁
k= heat dissipation factor	[]	2	2	1.2

$$\mathbf{P_r} = \frac{P_{\text{nom}} \times (t_1 + t_2) - (P_P \times t_1)}{t_2 \times k}$$


Practical Example (RACM600-48SL/OF for still air convection):


Take the RACM600-48SL/OF at 230VAC input Voltage and full load at $T_{\text{\tiny AMB}} = 50 \,^{\circ}\text{C}$, with still air convection.

$$P_P = 720W$$

=4s

$$P_r = \frac{380 \times (4 + 32) - (720 \times 4)}{32 \times 2} = \underline{169W}$$






SAFETY AND CERTIFICATIONS (DESIGNED TO MEET)		
Certificate Type (Safety)	Report Number	Standard
Audio/video, information and communication technology equipment. Safety requirements (CB)	T223-0662-21	IEC62368-1, 2nd Edition 2014
Audio/video, information and communication technology equipment. Safety requirements (LVD)	1223-0002-21	EN62368-1:2014 + A11:2017
Audio/Video, information and communication technology equipment -	E224736-A6026-	UL62368-1:2014
Part1: Safety requirements	UL	CAN/CSA-C22.2 No. 62368-1:2014
Medical Electric Equipment, General Requirements for Safety and Essential Performance	E314885-D1009-1/	ANSI/AAMI ES60601-1:2005A2:2010/(R)2012
Invedical Electric Equipment, denetal nequirements for Safety and Essential Performance	A0/C0-UL	CAN/CSA-C22.2 No. 60601-1:14, 3rd Edition
Medical electrical equipment Part 1: General requirements for basic safety and essential performance (CB Scheme)	T223-0661-21	IEC60601-1:2005, 3rd Edition + AM1:2012
Medical electrical equipment Part 1: General requirements for basic safety and essential performance	1 1223-0001-21	EN60601-1:2006 + A1:2013
RoHS2		RoHS 2011/65/EU



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

EMC Compliance	Condition	Standard / Criterion
Electromagnetic compatibility of multimedia equipment – Emission Requirements		EN55032:2015
Information technology equipment - Immunity characteristics - Limits and methods of measurement		EN55024:2010+A1:2015
Medical electrical equipment Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic disturbances — Requirements and tests		EN60601-1-2:2015
ESD Electrostatic Discharge Immunity Test	Air: ±15kV Contact: ±4,8kV	EN61000-4-2, Criteria A
Radiated, Radio-Frequency, Electromagnetic Field Immunity Test	level 3= 10V/m	EN61000-4-3, Criteria A
Fast Transient and Burst Immunity	level 4= ±4kV	EN61000-4-4, Criteria A
Surge Immunity	level 4= ±2kV DM, ±4kV CM	EN61000-4-5, Criteria A
Immunity to Conducted Disturbances, Induced by Radio-Frequency Fields	level= 3, 6Vrms in ISM band	EN61000-4-6, Criteria A
Power Magnetic Field Immunity	30A/m	EN61000-4-8, Criteria A
Voltage Dips	30%, 500ms 60%, 100ms 100%, 20ms	EN61000-4-11, Criteria A EN61000-4-11, Criteria B EN61000-4-11, Criteria A
Voltage Interruptions	30%, 500ms 60%, 100ms 100%, 20ms 100%, 5000ms	EN61000-4-11, Criteria A EN61000-4-11, Criteria B EN61000-4-11, Criteria A EN61000-4-11, Criteria B
Ring wave immunity test	level 3= 1kV DM, 2kV CM	EN61000-4-12, Class A
Voltage fluctuation immunity test for equipment with input current <16 A per phase	class 3	EN61000-4-14, Class A
Limits of Harmonic Current Emissions		EN61000-3-2:2014
Voltage Fluctuations and Flicker in Public Low-Voltage Systems		EN61000-3-3:2013




Parameter	Туре	Value
Material	case/baseplate	aluminum
Iviaterial	PCB	FR4
Dimension (LxWxH)		196.8 x 101.6 x 40.6mm
Weight		1000g typ



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)



Signal CON3					
	(Molex PCB Header 87833-1420)				
#	Function	#	Function		
14	-Sense	13	+Sense		
12	address	11	Current_share_V		
10	Remote ON/OFF	9	PSU_GOOD		
8	+5VSB	7	5VSB_RTN		
6	SDA	5	SCL		
4	5VSB_RTN	3	5VSB_RTN		
2	+5VSB	1	+5VSB		

Compatible Connector CON3		
Housing		
Molex 51110 Series or equivalent		
Crimp Terminal		
Molex 50394 Series or equivalent		

Input Terminal Block CON1 (9) (M3.5 screws) Dinkle: DT-4C-B01W-03-GN)		
Function	AWG	
PE	12-18	
L (line)	12-18	
N (neutral)	12-18	
wire stripping length: 7-8r	nm	

Note	es:				
	Note9: L	lse flexible	cable v	with bel	ow lugs:

recommended tightening torque: 1.3Nm

e9: Use flexible cable with below lugs:	
Output	
Input	

Output Terminal Block CON2 for 24 & 48Vout Version⁽⁹⁾ (M4 screws)

Dinkle: DT-7C-B01W-0137-02

Function	AWG
-VOUT	8-12
+VOUT	8-12
odne state also a less ette 40	44

wire stripping length: 10-11mm recommended tightening torque 1.5Nm

Output Terminal Block CON2 for 12Vout Version⁽⁹⁾

(M5 screws) Dinkle: 0166-80-S1531802C


Function	AWG
-VOUT	4-12
+VOUT	4-12

wire stripping length: 14-15mm recommended tightening torque 2.4Nm



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)






Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

MOUNTING INSTRUCTIONS

Mounting Orientations






horizontal (standard) mounting



With forced air cooling, mounting orientation has no impact on output power. Upside down mounting is not recommended. Forced air conditions as specified are valid for indicated airflow direction only (back side).

The PSU should be placed on a metal surface. It should not be placed on isolating and low thermal conductive surfaces. Take care that no objects can fall into the PSU.

Mounting Equipment



Recommended mounting tightening torque= 0.6Nm. Screw length= min. 2mm / max. 4mm

PACKAGING INFORMATION **Parameter** Type Value Packaging Dimension (LxWxH) cardboard box 400.0 x 318.0 x 150mm Packaging Quantity 7pcs -40°C to +85°C Storage Temperature Range Storage Humidity non-condensing 95% RH max.

The product information and specifications may be subject to changes even without prior written notice. The product has been designed for various applications; its suitability lies in the responsibility of each customer. The products are not authorized for use in safety-critical applications without RECOM's explicit written consent. A safety-critical application is an application where a failure may reasonably be expected to endanger or cause loss of life, inflict bodily harm or damage property. The applicant shall indemnify and hold harmless RECOM, its affiliated companies and its representatives against any damage claims in connection with the unauthorized use of RECOM products in such safety-critical applications.