Teledyne FLIR Commercial Systems, Inc.

10F, No. 57, Zhouzi Street, NeiHu District, Taipei city, 114, Taiwan

Declaration of Conformity

Model: ***DM266***

Description: ***Imaging Multimeter with IGM and Bluetooth***

Date of Issue: ***8/4/2026***

We, Teledyne FLIR Commercial Systems, Inc. 10F, No. 57, Zhouzi Street, NeiHu District, Taipei city, 114, Taiwan declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: **2014/30/EU**
Report Number: **260521177GZU-001**
Report Date of Issue: **6/24/2026**
Standards:
EN IEC 61326-1:2021
EN IEC 61326-2-2:2021

RED Directive **2014/53/EU**
RED Report # **260521177GZU-002; 260521177GZU-003; 260521177GZU-004**
Report Date of Issue: **6/24/2026**
Standards:
ETSI EN 301 489-1 V2.2.3 (2019-11)
ETSI EN 301 489-17 V3.3.1 (2024-09)
EN IEC 61000-3-2:2019+A1:2021+A2:2024
EN 61000-3-3:2013+A1:2019+A2:2021
EN 300328 V2.2.2 (2019-07)
EN 62479

LVD Directive: **2014/35/EU**
Report Number: **260205202GZU-001; 260205202GZU-002**
Report Date of Issue: **7/23/2026**
Standards:
UL61010-1:2012 Ed.3+R:15Nov2024
UL 61010-2-033:2020 Ed.2
CSA C22.2#61010-1-12:2012 Ed.3+U1;U2;A1;U3;U4
CSA C22.2#61010-2-033:2020 Ed.2
EN 61010-1:2010 +A1:2019
EN IEC 61010-2-033: 2021 +A11:2021

RoHS Directive: **2011/65/EU incl Am EU 2015/863**
Standard:
EN IEC 63000:2018

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive and EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive and EMC Directive.

TeledyneFLIR Commercial Systems

Hank Tsai

Hank Tsai
QA Engineering Mgmt.